Device Categories			Biological Effect Initial								Other ⁴		
Body Contact		Contact Duration A- Limited [≤ 24 hrs] B- Prolonged [>24 hrs to ≤30 days] C- Permanent [>30 days]	Cytotoxicity	Sensitization	Irritation	Systemic Toxicity	Subacute (Subchronic Toxicity	Genotoxicity	Implantation	Hemocompatibility	Chronic Toxicity	Carcinogenicity	
Surface Devices	Skin	A B C	•	•	•								
	Mucosal Membranes	A B C	•	•	•	♦	♦	-	 ◊ ◊ 		0		
	Breached or Compromised Surfaces	A B C	•			 ◊ ◊ ◊ ◊ 	<u>ہ</u>	-	 ◊ ◊ ◊ 		↓ ↓		
External Communicating Devices	Blood Path, Indirect ³ Tissue ¹ /Bone/Dentin Communicating	A B C A B C	8 8 8 8	•	• • •		 <	•	 <!--</td--><td></td><td>•</td><td></td>		•		
	Circulating Blood ³	A B C	•	•	•	•	•	0 ² ■	•	•	-	-	

ISO10993 Biocompatibility Testing of Medical Devices